



Potvrda o akreditaciji Accreditation Certificate

Ovime se utvrđuje da je

This is to recognize that

DIV Laboratorij d.o.o.

Inspekcijsko tijelo

Davora Zbiljskog 30, HR-10000 Zagreb

Inspekcijsko tijelo vrste A / Inspection Body Type A

osposobljen prema zahtjevima norme

is competent according to

HRN EN ISO/IEC 17020:2012

(ISO/IEC 17020:2012;

EN ISO/IEC 17020:2012)

za/to carry out

Inspekcija mjerila tlaka i opreme pod tlakom

Inspection of pressure gauges and pressure equipment

u području opisanom u prilogu koji je sastavni dio ove potvrde o akreditaciji.

for the scope described in the annex which is the constituent part of this accreditation certificate.

Br./No.: 6616

Klasa/Ref.No.: 383-02/18-75/003

Urbroj/Id.No.: 569-05/2-19-15

Zagreb, 2019-01-23

Akreditacija istječe•Accreditation expiry: 2024-01-22

Prva akreditacija•Initial accreditation: 2019-01-23

HAA je potpisnica multilateralnog sporazuma s Europskom organizacijom za akreditaciju (EA)

HAA is a signatory of the European co-operation for Accreditation (EA) Multilateral Agreement

Ravnatelj:

Director General:

Tihomir Babić, dipl. ing.



HAA

Hrvatska akreditacijska agencija

Croatian Accreditation Agency

PRILOG POTVRDI O AKREDITACIJI br: 6616

Annex to Accreditation Certificate No.

Klasa/Ref. No.: 383-02/18-75/003

Urbroj/Id. No.: 569-05/2-19-16

Datum izdanja priloga/Annex Issued on: 2019-01-23

Norma: HRN EN ISO/IEC 17020:2012

Standard: (ISO/IEC 17020:2012; EN ISO/IEC 17020:2012)

Akreditacija istječe: 2024-01-22

Accreditation expiry:

Prva akreditacija: 2019-01-23

Initial accreditation:

Akreditirano inspekcijsko tijelo vrste A:

Accredited Inspection Body Type A:

DIV Laboratorij d.o.o.

Inspekcijsko tijelo

Davora Zbiljskog 30, HR-10000Zagreb

Područje akreditacije:

Scope of Accreditation:

Inspekcija mjerila tlaka i opreme pod tlakom

Inspection of pressure gauges and pressure equipment

Važeće izdanje Priloga dostupno je na web adresi: www.akreditacija.hr /
Valid issue of the Annex is available at the web address: www.akreditacija.hr

Ravnatelj:

Director General:

Tihomir Babić, dipl. ing.

PODRUČJE AKREDITACIJE SCOPE OF ACCREDITATION

Br. No.	Predmet inspekcije <i>Field</i>	Vrsta inspekcije <i>Type and range</i>	Postupak inspekcije <i>Methods and procedures</i>
1.	<p>Mjerila tlaka (manometri, vakuummetri, tlakomjeri za gume, mjerni pretvornici tlaka) <i>Pressure gauges (manometers, vacuum meters, tyre pressure gauges, pressure transducers)</i></p>	<p>Prvi pregled <i>First inspection</i></p> <p>Redovna inspekcija <i>Regular inspection</i></p> <p>Izvanredna inspekcija <i>Exceptional inspection</i></p> <p>(-0,85 do 700) bar</p>	<p>Vlastiti postupak / Internal procedure IP-7-1, izdanje / Issue 01, 2017-10-02</p> <p>Temeljeno na / Based on:</p> <p>Naputak za ispitivanje pri ovjeravanju tlakomjera za gume (Narodne novine br. 11/2006, 13/2006) <i>Guide on testing of tyre pressure gauges for verification purpose (Official Gazette No. 11/2006, 13/2006)</i></p> <p>Naputak za umjeravanje radnih etalona manometara, vakuummetara, manovakuummetara i mjernih pretvornika tlaka (Narodne novine br. 11/2006) <i>Guide on testing of working standards of manometers, vacuum meters, mano-vacuum meters and pressure transducers for verification purpose (Official Gazette No. 11/2006)</i></p> <p>Uredba o posebnim uvjetima koje moraju ispunjavati ovlaštena tijela za obavljanje poslova ovjeravanja zakonitih mjerila i/ili poslova pripreme zakonitih mjerila za ovjeravanje (Narodne novine br. 90/2014, Dodatak 13) <i>Regulation on special requirements to be fulfilled by authorized bodies for performing verification of legal measuring instruments and/or activities of preparation of legal measuring instruments for verification (Official Gazette No. 90/2014, Annex 13)</i></p>
2.	<p>Oprema pod tlakom visoke razine opasnosti <i>Pressure equipment of high level of hazard</i></p>	<p>Inspekcija tijekom uporabe: <i>In-service inspection:</i></p> <p>Prvi pregled <i>First inspection</i></p> <p>Vanjski pregled <i>External inspection</i></p> <p>Unutarnji pregled <i>Internal inspection</i></p> <p>Ispitivanje tlakom <i>Pressure test</i></p> <p>Izvanredni pregled <i>Exceptional inspection</i></p> <p>Pregled prije ponovnog puštanja u rad <i>Inspection before restarting</i></p>	<p>Vlastiti postupak / Internal procedure: IP-7-2 izdanje / Issue 2, 2018-07-23</p> <p>Temeljeno na / Based on:</p> <p>Pravilnik o pregledima i ispitivanju opreme pod tlakom (Narodne novine br. 27/2017) <i>Ordinance on inspection and testing of pressure equipment (Official Gazette No. 27/2017)</i></p> <p>Tehničke upute i preporuke za provođenje aktivnosti periodičkih pregleda i ispitivanja opreme pod tlakom prema Pravilniku o pregledima i ispitivanju opreme pod tlakom (Narodne novine br. 27/2017), izdanje 2, 2017-10-18 <i>Technical instructions and recommendations for in service inspection and testing of pressure equipment according to Ordinance on inspection and testing of pressure equipment (Official Gazette No. 27/2017), Issue 2, 2017-10-18.</i> (https://www.mingo.hr/public/documents/tu_ppi_izdanje_2.pdf)</p>
3.	<p>Oprema pod tlakom niske razine opasnosti <i>Pressure equipment of low level of hazard</i></p>	<p>Dragovoljna inspekcija tijekom uporabe <i>Goodwill in-service inspection</i></p>	<p>Vlastiti postupak / Internal procedure: IP-7-2 izdanje / Issue 2, 2018-07-23</p> <p>Upute proizvođača <i>Manufacturer instructions</i></p>